

# Memo

**To:** Evelyn Foust  
Communicable Disease Branch Head

**From:** Angela Bradley, MLT, (ASCP)  
Laboratory Improvement

**CC:** Pete Moore  
Ron Higginbotham  
NCSLPH Lab Improvement Staff  
Local Health Department Lab Managers

**Date:** November 18th, 2016

**Re:** QC of GC-Lect Media

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

**GC-Lect:** Lot Number: **6300967** Expiration Date: **2/3/2017**  
Along with **Bio-Bag™ Type C:** Lot Number: **5286857** Expiration Date: **8/31/2017**  
and with **Bio-Bag™ Type C:** Lot Number: **6188936** Expiration Date: **4/30/2018**

### *Summary of NCSLPH QC results for this Lot #*

N. gonorrhoeae ATCC# 43069 - Growth	S. epidermidis ATCC# 12228 – No Growth
N. gonorrhoeae ATCC# 43070 - Growth	P. mirabilis ATCC# 43071 – No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

**Lot number failed:**  
None

BD and LABSCO® contacted: November 18th, 2016

\*If your contact information changes, please provide that information to Angie Bradley in Laboratory Improvement at (919) 733-7186 or via e-mail at [angela.bradley@dhhs.nc.gov](mailto:angela.bradley@dhhs.nc.gov). Thank you.

