



DEPARTMENT OF HEALTH AND HUMAN SERVICES
 DIVISION OF PUBLIC HEALTH, STATE LABORATORY OF PUBLIC HEALTH

ROY COOPER
 GOVERNOR

MANDY COHEN, MD, MPH
 SECRETARY

DANIEL STALEY
 DIRECTOR

SCOTT J. ZIMMERMAN, DRPH, MPH, HCLD (ABB)
 LABORATORY DIRECTOR

Memo

To: Evelyn Foust
 Communicable Disease Branch Head

From: Angela Bradley, MLT, (ASCP) Laboratory Improvement

CC: Pete Moore
 Ron Higginbotham
 NCSLPH Lab Improvement Staff
 Local Health Department Lab Managers

Date: March 17th, 2017

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

GC-Lect: Lot Number: **7055637** Expiration Date: **6/2/2017**
 Along with **Bio-Bag™ Type C:** Lot Number: **5286857** Expiration Date: **8/31/2017**
 and with **Bio-Bag™ Type C:** Lot Number: **6188936** Expiration Date: **4/30/2018**

Summary of NCSLPH QC results for this Lot #

| | |
|-------------------------------------|--|
| N. gonorrhoeae ATCC# 43069 - Growth | S. epidermidis ATCC# 12228 – No Growth |
| N. gonorrhoeae ATCC# 43070 - Growth | P. mirabilis ATCC# 43071 – No Growth |
| Sterility Check - Acceptable | Visual Check - Acceptable |

Lot number failed: None

BD and LABSCO® contacted: March 17th, 2017

*If your contact information changes, please provide that information to Angie Bradley in Laboratory Improvement at (919) 733-7186 or via e-mail at angela.bradley@dhhs.nc.gov. Thank you.