



North Carolina Department of Health and Human Services
State Laboratory of Public Health

Pat McCrory
Governor

Richard O. Brajer
Secretary

Scott J. Zimmerman, DrPH, MPH, HCLD (ABB)
Laboratory Director

November 20, 2015

MEMORANDUM

To: Evelyn Foust
Communicable Disease Branch Head

From: Tiffany Perdue, BS, MLS (ASCP)^{CM}
Laboratory Improvement Consultant

Subject: QC of GC-Lect Media

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are accepted for use:

GC-Lect: Lot Number: **5300915** Expiration Date: **2/5/2016**
Along with **Bio-Bag™ Type C:** Lot Number: **4155829** Expiration Date: **1/31/2016**

While performing quality control testing on the GC-Lect agar plates, there was a breakthrough growth of *Candida albicans* (yeast) at **72 hours**. Please pay attention to possible growth of other organisms, especially at the **72 hour** mark. Follow your laboratory's procedure for the presumptive identification of *Neisseria gonorrhoeae*.

BD and LabSco® contacted November 20, 2015

cc: Pete Moore
Ron Higginbotham
Patricia Atwood
LI Regional Consultants
Local Health Department Lab Managers
Lisa Ballance
Ann Brown
Angela Bradley
Diana Scarborough

*If your contact information changes, please provide that information to Tiffany Perdue in Laboratory Improvement at (919) 733-7186 or via e-mail at tiffany.perdue@dhhs.nc.gov. Thank you.

