

ROY COOPER • Governor

KODY H. KINSLEY • Secretary

MARK T. BENTON • Deputy Secretary for Health

SUSAN KANSAGRA MD, MBA Assistant • Secretary for Public Health Division of Public Health

SCOTT M. SHONE, PhD, HCLD (ABB) • Laboratory Director, State Laboratory of Public Health

Memo

To: Evelyn Foust

Communicable Disease Branch Head

From: Aja Stanback, BS

Quality Improvement Consultant

CC: Pete Moore

Ron Higginbotham

NCSLPH Lab Improvement Staff Local Health Department Lab Managers

Date: September 6, 2022

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

Modified Thayer Martin II: Lot Number: 146936P Expiration Date: 10/26/2022 Along with Bio-BagTM Type C: Lot Number: B03C342M Expiration Date: 9/30/2022 and with Bio-Bag TypeTM C: Lot Number: B020225M Expiration Date: 1/31/2023

Summary of NCSLPH QC results for this Lot #:

N. gonorrhoeae ATCC# 43069 - Growth	S. epidermidis ATCC# 12228 – No Growth
	*P. mirabilis ATCC# 43071 –1 colony at 48 hrs.
Sterility Check - Acceptable	Visual Check - Acceptable

Lot number failed: None

Hardy Diagnostics Modified Thayer Martin

*While performing quality control testing on Modified Thayer Martin II plates, there was breakthrough growth of *P. mirabilis* at **48 hours.** Please pay attention to possible growth of other organisms, especially at the **48-hour** mark. Follow your laboratory's procedure for the presumptive identification of *Neisseria gonorrhoeae*.

**If your contact information changes, please provide that information to Linda Cleve in Laboratory Improvement at (252) 414-3078 or via e-mail at Linda.cleve@dhhs.nc.gov. Thank you.

NC DEPARTMENT OF HEALTH AND HUMAN SERVICES \cdot DIVISION OF PUBLIC HEALTH \cdot STATE LABORATORY OF PUBLIC HEALTH