

ROY COOPER • Governor

MANDY COHEN, MD, MPH · Secretary

MARK T. BENTON • Assistant Secretary for Public Health

Division of Public Health

SCOTT M. SHONE, PhD, HCLD (ABB) • Laboratory Director, State Laboratory of Public Health

Memo

To: Evelyn Foust

Communicable Disease Branch Head

From: Angela Bradley, MLT (ASCP)

Laboratory Improvement Medical Laboratory Technologist

CC: Pete Moore

Ron Higginbotham

NCSLPH Lab Improvement Staff Local Health Department Lab Managers

Date: May 13th, 2021

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

Modified Thayer Martin II: Lot Number: 483529 Expiration Date: 07/01/2021 Along with Bio-BagTM Type C: Lot Number: B02C099M Expiration Date: 10/30/2021 and with Bio-BagTM Type C: Lot Number: B03C342M Expiration Date: 9/30/2022

Summary of NCSLPH QC results for this Lot #:

N. gonorrhoeae ATCC# 43069 - Growth	S. epidermidis ATCC# 12228 – No Growth
	P. mirabilis ATCC# 43071 –No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

Lot number failed: None

Hardy Diagnostics Modified Thayer Martin

**If your contact information changes, please provide that information to Angie Bradley in Laboratory Improvement at (919) 733-7186 or via e-mail at angela.bradley@dhhs.nc.gov. Thank you.