

ROY COOPER • Governor

MANDY COHEN, MD, MPH · Secretary

BETH LOVETTE, MPH, BSN, RN. Acting Director,

Division of Public Health

SCOTT J. ZIMMERMAN, DrPH, MPH, HCLD (ABB) • Laboratory Director, State Laboratory of Public Health

Memo

To: Evelyn Foust

Communicable Disease Branch Head

From: Angela Bradley, MLT (ASCP)

Medical Laboratory Technologist

CC: Pete Moore

Ron Higginbotham

NCSLPH Lab Improvement Staff Local Health Department Lab Managers

Date: <u>March 11th, 2019</u>

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

Modified Thayer Martin II: <u>Lot Number</u>: 9040775 Expiration Date: 6/07/2019 Along with Bio-Bag[™] Type C: <u>Lot Number</u>: 8129674 Expiration Date: 01/30/2020 and with Bio-Bag[™] Type C: Lot Number: 7296619 Expiration Date: 06/30/2019

Summary of NCSLPH QC results for this Lot #:

N. gonorrhoeae ATCC# 43069 - Growth	S. epidermidis ATCC# 12228 – No Growth
N. gonorrhoeae ATCC# 43070 - Growth	P. mirabilis ATCC# 43071 – * Growth
Sterility Check - Acceptable	Visual Check - Acceptable

^{*}While performing quality control testing on the Modified Thayer Martin II agar plates, there was growth of 1 colony of P. mirabilis (Oxidase-negative, Gram negative rod) at 48 hours. Please pay attention to possible growth of other organisms, especially at the 48-hour mark. Follow your laboratory's procedure for the presumptive identification of Neisseria gonorrhoeae.

Lot number failed: None

BD and Fisher Scientific contacted: March 8th, 2019.

*If your contact information changes, please provide that information to Angie Bradley in Laboratory Improvement at (919) 733-7186 or via e-mail at angela.bradley@dhhs.nc.gov. Thank you.

NC DEPARTMENT OF HEALTH AND HUMAN SERVICES • DIVISION OF PUBLIC HEALTH • STATE LABORATORY OF PUBLIC HEALTH