

**ROY COOPER • Governor** 

KODY H. KINSLEY · Secretary

MARK BENTON • Deputy Secretary for Health

**SUSAN KANSAGRA MD, MBA •** Assistant Secretary for Public Health Division of Public Health

**SCOTT M. SHONE, PhD, HCLD (ABB)** • Laboratory Director, State Laboratory of Public Health

## Memo

**To:** Evelyn Foust

Communicable Disease Branch Head

From: Lettia Spruill

Medical Laboratory Technologist I

**CC:** Pete Moore

Ron Higginbotham

NCSLPH Lab Improvement Staff Local Health Department Lab Managers

## Date:

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

Modified Thayer Martin II: Lot Number: 611710P Expiration Date: 12/06/2023 and Modified Thayer Martin II: Lot Number: 611081P Expiration Date: 12/01/2023

Along with Bio-Bag™ Type C: Lot Number: B01E182M Expiration Date: 02/28/2024

## **Summary of NCSLPH QC results for this Lot #:**

N. gonorrhoeae ATCC# 43069 - Growth	S. epidermidis ATCC# 12228 – No Growth
	P. mirabilis ATCC# 43071 – No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

**Lot number failed:** None

Hardy Diagnostics Modified Thayer Martin

\*\*If your contact information changes, please provide that information to Tracey Shives in Laboratory Improvement at (336)-306-4302 or via email at tracey.shives@dhhs.nc.gov. Thank you.

NC DEPARTMENT OF HEALTH AND HUMAN SERVICES • DIVISION OF PUBLIC HEALTH • STATE LABORATORY OF PUBLIC HEALTH