

# Memo

**To:** Evelyn Foust  
Communicable Disease Branch Head

**From:** Angela Bradley, MLT, (ASCP)  
Laboratory Improvement

**CC:** Pete Moore  
Ron Higginbotham  
NCSLPH Lab Improvement Staff  
Local Health Department Lab Managers

**Date:** October 19th, 2016

**Re:** QC of BioBag

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

**Bio-Bag™ Type C:** Lot Number: **6188936** Expiration Date: **4/30/2018**

### *Summary of NCSLPH QC results for this Lot #*

N. gonorrhoeae ATCC# 43069 - Growth
N. gonorrhoeae ATCC# 43070 - Growth

Lot number failed:  
None

BD and LabSco® contacted: October 14th, 2016

\*If your contact information changes, please provide that information to Angie Bradley in Laboratory Improvement at (919) 733-7186 or via e-mail at [angela.bradley@dhhs.nc.gov](mailto:angela.bradley@dhhs.nc.gov). Thank you.

