

Memo

To: Evelyn Foust
Communicable Disease Branch Head

From: Angela Bradley, MLT, (ASCP)
Laboratory Improvement

CC: Pete Moore
Ron Higginbotham
NCSLPH Lab Improvement Staff
Local Health Department Lab Managers

Date: November 4th, 2016

Re: QC of BioBag

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

Bio-Bag™ Type C: Lot Number: 6257773 Expiration Date: 7/31/2018

Summary of NCSLPH QC results for this Lot #

N. gonorrhoeae ATCC# 43069 - Growth
N. gonorrhoeae ATCC# 43070 - Growth

Lot number failed:
None

BD and LABSCO® contacted: October 28th, 2016

*If your contact information changes, please provide that information to Angie Bradley in Laboratory Improvement at (919) 733-7186 or via e-mail at angela.bradley@dhhs.nc.gov. Thank you.

