

Memo

To: Evelyn Foust
Communicable Disease Branch Head

From: Angela Bradley, MLT, (ASCP)
Laboratory Improvement

CC: Pete Moore
Ron Higginbotham
NCSLPH Lab Improvement Staff
Local Health Department Lab Managers

Date: September 6th, 2016

Re: QC of GC-Lect Media

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

GC-Lect: Lot Number: **6223673** Expiration Date: **11/18/2016**
Along with **Bio-Bag™ Type C:** Lot Number: **5286857** Expiration Date: **8/31/2017**

Summary of NCSLPH QC results for this Lot #

N. gonorrhoeae ATCC# 43069 - Growth	S. epidermidis ATCC# 12228 – No Growth
N. gonorrhoeae ATCC# 43070 - Growth	P. mirabilis ATCC# 43071 – No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

Lot number failed:
None

BD and LabSco® contacted: September 6th, 2016

*If your contact information changes, please provide that information to Angie Bradley in Laboratory Improvement at (919) 733-7186 or via e-mail at angela.bradley@dhhs.nc.gov. Thank you.

