



DEPARTMENT OF HEALTH AND HUMAN SERVICES
 DIVISION OF PUBLIC HEALTH, STATE LABORATORY OF PUBLIC HEALTH

ROY COOPER
 GOVERNOR

MANDY COHEN, MD, MPH
 SECRETARY

DANIEL STALEY
 DIRECTOR

SCOTT J. ZIMMERMAN, DRPH, MPH, HCLD (ABB)
 LABORATORY DIRECTOR

Memo

To: Evelyn Foust
 Communicable Disease Branch Head

From: Heather Colvin, B.S.
 Laboratory Improvement Consultant

CC: Pete Moore
 Ron Higginbotham
 NCSLPH Lab Improvement Staff
 Local Health Department Lab Managers

Date: June 23rd, 2017

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

GC-Lect: Lot Number: **7152881** Expiration Date: **09/08/2017**
 Along with **Bio-Bag™ Type C:** Lot Number: **6188936** Expiration Date: **04/30/2018**
 and with **Bio-Bag™ Type C:** Lot Number: **6257773** Expiration Date: **07/31/2018**

Summary of NCSLPH QC results for this Lot #:

N. gonorrhoeae ATCC# 43069 - Growth	S. epidermidis ATCC# 12228 – No Growth
N. gonorrhoeae ATCC# 43070 - Growth	P. mirabilis ATCC# 43071 – No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

Lot number failed: None

BD and LABSCO® contacted: June 23rd, 2017

*If your contact information changes, please provide that information to Heather Colvin in Laboratory Improvement at (919) 733-7186 or via e-mail at heather.colvin@dhhs.nc.gov, Thank you!

HTTP://SLPH.NCPUBLICHEALTH.COM
 TEL 919-733-7834 • FAX 919-733-8695
 LOCATION: 4312 DISTRICT DRIVE • RALEIGH, NC 27607
 MAILING ADDRESS: 1918 MAIL SERVICE CENTER • RALEIGH, NC 27699-1918
 P.O. Box 28047 • RALEIGH, NC 27611-8047
 AN EQUAL OPPORTUNITY / AFFIRMATIVE ACTION EMPLOYER