



DEPARTMENT OF HEALTH AND HUMAN SERVICES
DIVISION OF PUBLIC HEALTH, STATE LABORATORY OF PUBLIC HEALTH

ROY COOPER
GOVERNOR

MANDY COHEN, MD, MPH
SECRETARY

DANIEL STALEY
DIRECTOR

SCOTT J. ZIMMERMAN, DRPH, MPH, HCLD (ABB)
LABORATORY DIRECTOR

Memo

To: Evelyn Foust
Communicable Disease Branch Head

From: Angela Bradley, MLT, ASCP
Laboratory Technologist

CC: Pete Moore
Ron Higginbotham
NCSLPH Lab Improvement Staff
Local Health Department Lab Managers

Date: July 16th, 2018

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

GC-Lect: Lot Number: **8170987** Expiration Date: **09/21/2018**
Along with **Bio-Bag™ Type C:** Lot Number: **7128668** Expiration Date: **02/28/2019**
and with **Bio-Bag™ Type C:** Lot Number: **7296619** Expiration Date: **06/30/2019**

Summary of NCSLPH QC results for this Lot #:

<i>N. gonorrhoeae</i> ATCC# 43069 - Growth	<i>S. epidermidis</i> ATCC# 12228 – No Growth
<i>N. gonorrhoeae</i> ATCC# 43070 - Growth	<i>P. mirabilis</i> ATCC# 43071 – No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

Lot number failed: None

BD and Fisher Scientific contacted: July 13th, 2018.

*If your contact information changes, please provide that information to Angie Bradley in Laboratory Improvement at (919) 733-7186 or via e-mail at angela.bradley@dhhs.nc.gov. Thank you.