

Memo

To: Evelyn Foust
Communicable Disease Branch Head

From: Angela Bradley, MLT, (ASCP)
Laboratory Improvement

CC: Pete Moore
Ron Higginbotham
NCSLPH Lab Improvement Staff
Local Health Department Lab Managers

Date: August 5th, 2016

Re: QC of GC-Lect Media

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

GC-Lect: Lot Number: **6196818** Expiration Date: **10/21/2016**
Along with **Bio-Bag™ Type C:** Lot Number: **5286857** Expiration Date: **8/31/2017**

Lot number failed:
None

BD and LabSco® contacted: August 5, 2016

*If your contact information changes, please provide that information to Angie Bradley in Laboratory Improvement at (919) 733-7186 or via e-mail at angela.bradley@dhhs.nc.gov. Thank you.